

Intravital Microscopy of Microfoam Injected into Rat Cremaster Microvasculature

Page 9


Air-Based Polidocanol Foam (Tessari Method)

Still photo of air-based foam
bubbles lodged in arterioles; no
change over 30 minutes



"OLD" Microfoam 7% nitrogen

Video-frame capture of
proprietary microbubbles which
were much smaller and did not
obstruct arterioles



"NEW" Microfoam 0.01-0.8% nitrogen

Video-frame capture of bubbles
only visible with image
enhancement; did not obstruct
arterioles

